Ohio Department of Health

Alcohol and Drug Testing

Inspector's Certification Statement

I Craig Yanni am a representative of the Director of Health. On 7-26-16 I performed a certification on the Intoxilyzer model 8000 (OH-5) with the serial number 80-00 4607. This certification is within the acceptable time frame as set forth by O.A.C. 3701-53-04(C).

The Instrument performed a dry gas control test before and after the instrument certification using a dry gas standard traceable to the National Institute of Standards and Technology (NIST). The dry gas control results are valid and are at or within five one-thousandths (+/- 0.005) grams per two hundred ten liters of the alcohol concentration on the manufacturer's certificate of analysis for that dry gas standard.

I performed the certification on this approved evidential breath testing instrument using a solution containing ethyl alcohol approved by the director of health according to the instrument display for the instrument being certified. The certification was valid and the certification results were at or within five one-thousands (+/- 0.005) grams per two hundred ten liters of the target value for that approved solution.

On the date I performed the certification, I opened a new bottle of solution. When I was done using the solution I disposed of the solution and the bottle, thus there was no need to refrigerate the solution. The bottle and solution was discarded the same day it was open and thus not retained for future reference. This bottle was not used more than three months after its date of first use, or after the manufacturer's expiration date on the approved solution certificate.

Certification results and all certification paperwork are retained in a manner prescribed by the Director of Health. The original documents are scanned and placed on the internet web site, corresponding with the appropriate serial number. The original documents are then filed and kept in the ordinary course of business in the office of Alcohol and Drug Testing. These documents will be kept for at least 3 years, in compliance with O.A.C. 3701-53-01(A).

CSYanni — 7-26-16

DIRECTORS REPRESENTATIVE'S SIGNATURE — DATE